

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY102521\  
 Data File : VY006471.D  
 Acq On : 25 Oct 2021 17:26  
 Operator : SY/MD  
 Sample : M4290-10RE  
 Misc : 4.80G/5ML/MSVOA\_Y/SOIL  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 TH-BBSRE

Quant Time: Oct 26 02:45:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y102221S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 23 06:59:04 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	7.795	168	73866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	124505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	101591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	44000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	53228	58.559	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.120%
35) Dibromofluoromethane	7.716	113	37599	54.181	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.360%
50) Toluene-d8	10.179	98	141868	48.225	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.460%
62) 4-Bromofluorobenzene	12.483	95	40833	41.205	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.420%
Target Compounds						
20) Methylene Chloride	4.716	84	12596	12.480	ug/l	91
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY102521\  
 Data File : VY006471.D  
 Acq On : 25 Oct 2021 17:26  
 Operator : SY/MD  
 Sample : M4290-10RE  
 Misc : 4.80G/5ML/MSVOA\_Y/SOIL  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 TH-BBSRE

Quant Time: Oct 26 02:45:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y102221S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 23 06:59:04 2021  
 Response via : Initial Calibration

