



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
 Data File : VY006453.D
 Acq On : 25 Oct 2021 07:48
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y1022215.M
 Title : SW846 8260
 Last Update : Sat Oct 23 06:59:04 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.2	20397	PASS
75	95	30	60	57.1	50216	PASS
95	95	100	100	100.0	87987	PASS
96	95	5	9	6.3	5539	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.9	77373	PASS
175	174	5	9	8.2	6323	PASS
176	174	95	101	95.3	73715	PASS
177	176	5	9	6.8	4980	PASS