

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110122\
 Data File : VY011221.D
 Acq On : 01 Nov 2022 14:57
 Operator : KP/MD
 Sample : N5316-06RE
 Misc : 8.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-GRAB-6-NE-QUARTERRE

Quant Time: Nov 01 15:11:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	7479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	11405	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	9460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	4705	58.821	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.640%
35) Dibromofluoromethane	7.709	113	2732	37.283	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.560%
50) Toluene-d8	10.179	98	14066	49.441	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.477	95	3719	37.654	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110122\
Data File : VY011221.D
Acq On : 01 Nov 2022 14:57
Operator : KP/MD
Sample : N5316-06RE
Misc : 8.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-GRAB-6-NE-QUARTERRE

Quant Time: Nov 01 15:11:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

