

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110122\
 Data File : VY011226.D
 Acq On : 01 Nov 2022 16:50
 Operator : KP/MD
 Sample : N5374-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-2-102822

Quant Time: Nov 02 01:43:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	219455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	376766	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	329666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	146515	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	140082	59.684	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.360%
35) Dibromofluoromethane	7.710	113	148398	61.304	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.600%
50) Toluene-d8	10.179	98	510632	54.331	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.660%
62) 4-Bromofluorobenzene	12.477	95	165162	50.619	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110122\
Data File : VY011226.D
Acq On : 01 Nov 2022 16:50
Operator : KP/MD
Sample : N5374-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-2-102822

Quant Time: Nov 02 01:43:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

