

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110122\
 Data File : VY011239.D
 Acq On : 01 Nov 2022 23:15
 Operator : KP/MD
 Sample : N5395-07
 Misc : 0.14g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-3-103122

Quant Time: Nov 02 01:45:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	203103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	368632	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	333862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	147143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	138969	63.976	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.960%	
35) Dibromofluoromethane	7.716	113	138094	58.306	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.620%	
50) Toluene-d8	10.179	98	486927	52.952	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.900%	
62) 4-Bromofluorobenzene	12.477	95	160276	50.206	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.420%	
Target Compounds						Qvalue
16) Acetone	3.942	43	9696	14.787	ug/l #	75
20) Methylene Chloride	4.710	84	61453	19.888	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110122\
Data File : VY011239.D
Acq On : 01 Nov 2022 23:15
Operator : KP/MD
Sample : N5395-07
Misc : 0.14g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-3-103122

Quant Time: Nov 02 01:45:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

