

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110122\  
 Data File : VY011240.D  
 Acq On : 01 Nov 2022 23:38  
 Operator : KP/MD  
 Sample : N5336-01  
 Misc : 5.05g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 SB-01-0-2.0

Quant Time: Nov 02 01:45:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y110122S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 01 11:56:18 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	7.789	168	10654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	19160	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	18914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	7635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	9646	84.655	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 169.300%#	
35) Dibromofluoromethane	7.710	113	6787	55.133	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.260%	
50) Toluene-d8	10.179	98	22198	46.444	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.880%	
62) 4-Bromofluorobenzene	12.483	95	8111	48.883	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 97.760%	
Target Compounds						
					Qvalue	
10) Methyl Iodide	4.101	142	1448	9.597	ug/l #	83
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110122\  
 Data File : VY011240.D  
 Acq On : 01 Nov 2022 23:38  
 Operator : KP/MD  
 Sample : N5336-01  
 Misc : 5.05g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 SB-01-0-2.0

Quant Time: Nov 02 01:45:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y110122S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 01 11:56:18 2022  
 Response via : Initial Calibration

