

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110122\  
 Data File : VY011217.D  
 Acq On : 01 Nov 2022 13:01  
 Operator : KP/MD  
 Sample : VSTDICV050  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 ICVVY110122

Manual Integrations  
 APPROVED

Reviewed By :Krupa  
 Patel

11/02/2022  
 Supervised By :Mahesh  
 Dadoda

Quant Time: Nov 01 14:32:20 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y110122S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 01 11:56:18 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 7.783  | 168  | 265366   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.685  | 114  | 424052   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.489 | 117  | 384968   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.422 | 152  | 186299   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 8.136  | 65    | 132410   | 46.654   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | =    | 93.300% |
| 35) Dibromofluoromethane    | 7.716  | 113   | 133675   | 49.064   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | =    | 98.120% |
| 50) Toluene-d8              | 10.179 | 98    | 512452   | 48.444   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 49 - 140 | Recovery | =    | 96.880% |
| 62) 4-Bromofluorobenzene    | 12.477 | 95    | 180314   | 49.101   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 25 - 144 | Recovery | =    | 98.200% |

|                              |       |     |         |         |        |     |
|------------------------------|-------|-----|---------|---------|--------|-----|
| Target Compounds             |       |     |         |         | Qvalue |     |
| 2) Dichlorodifluoromethane   | 1.906 | 85  | 102150  | 48.594  | ug/l   | 100 |
| 3) Chloromethane             | 2.113 | 50  | 138996  | 47.237  | ug/l   | 95  |
| 4) Vinyl Chloride            | 2.247 | 62  | 155745  | 47.383  | ug/l   | 98  |
| 5) Bromomethane              | 2.643 | 94  | 107273  | 48.919  | ug/l   | 99  |
| 6) Chloroethane              | 2.790 | 64  | 101226  | 47.926  | ug/l   | 100 |
| 7) Trichlorofluoromethane    | 3.119 | 101 | 236179  | 48.637  | ug/l   | 95  |
| 8) Diethyl Ether             | 3.521 | 74  | 87529   | 50.234  | ug/l   | 80  |
| 9) 1,1,2-Trichlorotrifluo... | 3.893 | 101 | 146053  | 50.026  | ug/l   | 94  |
| 10) Methyl Iodide            | 4.088 | 142 | 186613  | 49.655  | ug/l   | 91  |
| 11) Tert butyl alcohol       | 4.948 | 59  | 79942   | 276.264 | ug/l   | 99  |
| 12) 1,1-Dichloroethene       | 3.869 | 96  | 144731  | 50.140  | ug/l   | 82  |
| 13) Acrolein                 | 3.723 | 56  | 33792   | 220.660 | ug/l   | 98  |
| 14) Allyl chloride           | 4.472 | 41  | 223332  | 51.078  | ug/l # | 84  |
| 15) Acrylonitrile            | 5.155 | 53  | 214544  | 245.380 | ug/l   | 96  |
| 16) Acetone                  | 3.942 | 43  | 235624  | 275.032 | ug/l # | 84  |
| 17) Carbon Disulfide         | 4.186 | 76  | 467385  | 49.298  | ug/l   | 100 |
| 18) Methyl Acetate           | 4.472 | 43  | 102622  | 45.245  | ug/l # | 87  |
| 19) Methyl tert-butyl Ether  | 5.216 | 73  | 389740  | 50.810  | ug/l   | 96  |
| 20) Methylene Chloride       | 4.704 | 84  | 179174  | 51.267  | ug/l # | 82  |
| 21) trans-1,2-Dichloroethene | 5.216 | 96  | 165265  | 50.338  | ug/l   | 88  |
| 22) Diisopropyl ether        | 6.112 | 45  | 465729  | 51.375  | ug/l # | 90  |
| 23) Vinyl Acetate            | 6.051 | 43  | 1438678 | 260.013 | ug/l # | 91  |
| 24) 1,1-Dichloroethane       | 6.009 | 63  | 282772  | 49.809  | ug/l   | 99  |
| 25) 2-Butanone               | 6.978 | 43  | 302630  | 251.055 | ug/l # | 84  |
| 26) 2,2-Dichloropropane      | 6.972 | 77  | 252778  | 48.695  | ug/l   | 95  |
| 27) cis-1,2-Dichloroethene   | 6.978 | 96  | 183774  | 50.801  | ug/l   | 88  |
| 28) Bromochloromethane       | 7.326 | 49  | 109694  | 50.302  | ug/l   | 86  |
| 29) Tetrahydrofuran          | 7.344 | 42  | 173541  | 247.295 | ug/l # | 84  |
| 30) Chloroform               | 7.502 | 83  | 284507  | 49.620  | ug/l   | 98  |
| 31) Cyclohexane              | 7.777 | 56  | 262865  | 49.971  | ug/l   | 95  |
| 32) 1,1,1-Trichloroethane    | 7.697 | 97  | 251796  | 49.511  | ug/l   | 95  |
| 36) 1,1-Dichloropropene      | 7.911 | 75  | 227798  | 51.199  | ug/l   | 98  |
| 37) Ethyl Acetate            | 7.069 | 43  | 119120  | 49.665  | ug/l # | 93  |
| 38) Carbon Tetrachloride     | 7.899 | 117 | 233031  | 50.484  | ug/l   | 98  |
| 39) Methylcyclohexane        | 9.179 | 83  | 288644  | 53.325  | ug/l   | 91  |
| 40) Benzene                  | 8.155 | 78  | 658976  | 50.394  | ug/l   | 96  |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110122\  
 Data File : VY011217.D  
 Acq On : 01 Nov 2022 13:01  
 Operator : KP/MD  
 Sample : VSTDICV050  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 ICVVY110122

Manual Integrations  
 APPROVED

Reviewed By :Krupa  
 Patel

Quant Time: Nov 01 14:32:20 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y110122S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 01 11:56:18 2022  
 Response via : Initial Calibration

11/02/2022  
 Supervised By :Mahesh  
 Dadoda

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 41) Methacrylonitrile         | 7.301  | 41   | 69703m   | 53.484   | ug/l   |          |
| 42) 1,2-Dichloroethane        | 8.234  | 62   | 175820   | 49.514   | ug/l   | 90       |
| 43) Isopropyl Acetate         | 8.270  | 43   | 211883   | 50.141   | ug/l # | 88       |
| 44) Trichloroethene           | 8.935  | 130  | 175822   | 50.242   | ug/l   | 95       |
| 45) 1,2-Dichloropropane       | 9.209  | 63   | 161673   | 50.010   | ug/l   | 95       |
| 46) Dibromomethane            | 9.301  | 93   | 90915    | 49.773   | ug/l   | 96       |
| 47) Bromodichloromethane      | 9.496  | 83   | 218150   | 50.381   | ug/l   | 99       |
| 48) Methyl methacrylate       | 9.289  | 41   | 100180   | 51.613   | ug/l # | 78       |
| 49) 1,4-Dioxane               | 9.295  | 88   | 26183    | 1065.519 | ug/l # | 82       |
| 51) 4-Methyl-2-Pentanone      | 10.069 | 43   | 578929   | 251.963  | ug/l # | 86       |
| 52) Toluene                   | 10.240 | 92   | 415691   | 51.268   | ug/l   | 95       |
| 53) t-1,3-Dichloropropene     | 10.465 | 75   | 231495   | 50.765   | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 9.923  | 75   | 266619   | 51.097   | ug/l # | 82       |
| 55) 1,1,2-Trichloroethane     | 10.642 | 97   | 129010   | 49.280   | ug/l   | 98       |
| 56) Ethyl methacrylate        | 10.508 | 69   | 182405   | 52.749   | ug/l # | 76       |
| 57) 1,3-Dichloropropane       | 10.788 | 76   | 222432   | 49.639   | ug/l   | 97       |
| 58) 2-Chloroethyl Vinyl ether | 9.782  | 63   | 415036   | 256.018  | ug/l   | 96       |
| 59) 2-Hexanone                | 10.831 | 43   | 430688   | 262.555  | ug/l   | 83       |
| 60) Dibromochloromethane      | 10.983 | 129  | 154367   | 49.629   | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 11.087 | 107  | 123429   | 49.662   | ug/l   | 100      |
| 64) Tetrachloroethene         | 10.715 | 164  | 166244   | 50.838   | ug/l   | 97       |
| 65) Chlorobenzene             | 11.514 | 112  | 435508   | 51.277   | ug/l   | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 11.587 | 131  | 159386   | 50.967   | ug/l   | 98       |
| 67) Ethyl Benzene             | 11.593 | 91   | 796692   | 52.881   | ug/l   | 99       |
| 68) m/p-Xylenes               | 11.703 | 106  | 615421   | 105.232  | ug/l   | 94       |
| 69) o-Xylene                  | 12.026 | 106  | 289561   | 52.820   | ug/l   | 94       |
| 70) Styrene                   | 12.044 | 104  | 489531   | 52.675   | ug/l   | 96       |
| 71) Bromoform                 | 12.209 | 173  | 99002    | 50.935   | ug/l # | 98       |
| 73) Isopropylbenzene          | 12.331 | 105  | 760306   | 54.585   | ug/l   | 100      |
| 74) N-amyl acetate            | 12.142 | 43   | 186864   | 53.444   | ug/l # | 84       |
| 75) 1,1,2,2-Tetrachloroethane | 12.581 | 83   | 147458   | 51.159   | ug/l   | 99       |
| 76) 1,2,3-Trichloropropane    | 12.629 | 75   | 109891m  | 48.452   | ug/l   |          |
| 77) Bromobenzene              | 12.605 | 156  | 173980   | 53.101   | ug/l   | 98       |
| 78) n-propylbenzene           | 12.672 | 91   | 928169   | 54.457   | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 12.757 | 91   | 519416   | 53.187   | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 12.812 | 105  | 642019   | 54.470   | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 12.380 | 75   | 54541    | 51.465   | ug/l # | 83       |
| 82) 4-Chlorotoluene           | 12.855 | 91   | 535006   | 53.161   | ug/l   | 100      |
| 83) tert-Butylbenzene         | 13.075 | 119  | 552304   | 54.382   | ug/l   | 94       |
| 84) 1,2,4-Trimethylbenzene    | 13.117 | 105  | 633862   | 54.374   | ug/l   | 100      |
| 85) sec-Butylbenzene          | 13.251 | 105  | 821835   | 54.578   | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 13.367 | 119  | 688525   | 54.859   | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 13.367 | 146  | 341961   | 52.051   | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 13.446 | 146  | 336381   | 51.052   | ug/l   | 98       |
| 89) n-Butylbenzene            | 13.696 | 91   | 641612   | 54.745   | ug/l   | 99       |
| 90) Hexachloroethane          | 13.958 | 117  | 129861   | 51.736   | ug/l   | 90       |
| 91) 1,2-Dichlorobenzene       | 13.739 | 146  | 304353   | 51.864   | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.355 | 75   | 24418    | 49.453   | ug/l   | 83       |
| 93) 1,2,4-Trichlorobenzene    | 15.007 | 180  | 184060   | 51.894   | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 15.111 | 225  | 106745   | 51.687   | ug/l   | 99       |
| 95) Naphthalene               | 15.233 | 128  | 381439   | 46.312   | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.422 | 180  | 158295   | 50.784   | ug/l   | 99       |

11/02/2022

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110122\  
Data File : VY011217.D  
Acq On : 01 Nov 2022 13:01  
Operator : KP/MD  
Sample : VSTDICV050  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Nov 01 14:32:20 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y110122S.M  
Quant Title : SW846 8260  
QLast Update : Tue Nov 01 11:56:18 2022  
Response via : Initial Calibration

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda

11/02/2022

| Compound                                                                    | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------------------------------------------------------|------|------|----------|------|-------|----------|
| -----                                                                       |      |      |          |      |       |          |
| (# ) = qualifier out of range (m) = manual integration (+) = signals summed |      |      |          |      |       |          |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110122\  
 Data File : VY011217.D  
 Acq On : 01 Nov 2022 13:01  
 Operator : KP/MD  
 Sample : VSTDICV050  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 ICVVY110122


Quant Time: Nov 01 14:32:20 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y110122S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 01 11:56:18 2022  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By :Krupa  
 Patel

11/02/2022  
 Supervised By :Mahesh  
 Dadoda


11/02/2022





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.783 min Scan# 91  
Delta R.T. -0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:168 Resp: 265360  
Ion Ratio Lower Upper  
168 100  
99 55.7 46.9 70.3

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#2  
Dichlorodifluoromethane  
Concen: 48.594 ug/l  
RT: 1.906 min Scan# 14  
Delta R.T. 0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 85 Resp: 102150  
Ion Ratio Lower Upper  
85 100  
87 32.2 16.1 48.2





#3  
Chloromethane  
Concen: 47.237 ug/l  
RT: 2.113 min Scan# 48  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 50 Resp: 138990  
Ion Ratio Lower Upper  
50 100  
52 33.6 24.6 37.0

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#4  
Vinyl Chloride  
Concen: 47.383 ug/l  
RT: 2.247 min Scan# 70  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 62 Resp: 155745  
Ion Ratio Lower Upper  
62 100  
64 32.4 25.0 37.6





#5  
Bromomethane  
Concen: 48.919 ug/l  
RT: 2.643 min Scan# 11  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 94 Resp: 10727  
Ion Ratio Lower Upper  
94 100  
96 92.7 73.5 110.3

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#6  
Chloroethane  
Concen: 47.926 ug/l  
RT: 2.790 min Scan# 159  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 64 Resp: 101226  
Ion Ratio Lower Upper  
64 100  
66 31.8 25.4 38.0





#7  
Trichlorofluoromethane  
Concen: 48.637 ug/l  
RT: 3.119 min Scan# 213  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:101 Resp: 236179  
Ion Ratio Lower Upper  
101 100  
103 63.5 48.0 72.0

Manual Integrations  
APPROVED


Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda



11/02/2022




#8  
Diethyl Ether  
Concen: 50.234 ug/l  
RT: 3.521 min Scan# 279  
Delta R.T. -0.007 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01



Tgt Ion: 74 Resp: 87529  
Ion Ratio Lower Upper  
74 100  
45 89.1 55.1 165.5





#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 50.026 ug/l  
RT: 3.893 min Scan# 340  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122





Tgt Ion:101 Resp: 14605  
Ion Ratio Lower Upper  
101 100  
85 43.6 40.6 60.8  
151 78.1 64.9 97.3

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#10  
Methyl Iodide  
Concen: 49.655 ug/l  
RT: 4.088 min Scan# 372  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion:142 Resp: 186613  
Ion Ratio Lower Upper  
142 100  
127 43.7 41.5 62.3  
141 14.4 11.9 17.9





#11

Tert butyl alcohol

Concen: 276.264 ug/l

RT: 4.948 min Scan# 514

Delta R.T. -0.006 min

Lab File: VY011217.D




Acq: 01 Nov 2022 13:01

Tgt Ion: 59 Resp: 7994

Ion Ratio Lower Upper

59 100

57 8.4 6.4 9.6



Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122

Manual Integrations


APPROVED

Reviewed By :Krupa  
Patel

11/02/2022

Supervised By :Mahesh  
Dadoda

11/02/2022



#12

1,1-Dichloroethene

Concen: 50.140 ug/l

RT: 3.869 min Scan# 336

Delta R.T. -0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01




Tgt Ion: 96 Resp: 144731


Ion Ratio Lower Upper

96 100

61 149.5 145.8 218.6


98 63.2 52.5 78.7





#13  
Acrolein  
Concen: 220.660 ug/l  
RT: 3.723 min Scan# 312  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122





Tgt Ion: 56 Resp: 33792  
Ion Ratio Lower Upper  
56 100  
55 70.0 57.6 86.4

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#14  
Allyl chloride  
Concen: 51.078 ug/l  
RT: 4.472 min Scan# 435  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 41 Resp: 223332  
Ion Ratio Lower Upper  
41 100  
39 61.3 58.8 88.2  
76 41.5 25.1 37.7#





#15  
Acrylonitrile  
Concen: 245.380 ug/l  
RT: 5.155 min Scan# 547  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 53 Resp: 214544  
Ion Ratio Lower Upper  
53 100  
52 80.5 67.0 100.4  
51 35.0 29.8 44.6

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#16  
Acetone  
Concen: 275.032 ug/l  
RT: 3.942 min Scan# 348  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 43 Resp: 235624  
Ion Ratio Lower Upper  
43 100  
58 35.9 22.0 33.0#





#17  
Carbon Disulfide  
Concen: 49.298 ug/l  
RT: 4.186 min Scan# 388  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 76 Resp: 467385  
Ion Ratio Lower Upper  
76 100  
78 9.1 7.4 11.0

Manual Integrations  
APPROVED


Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda



11/02/2022




#18  
Methyl Acetate  
Concen: 45.245 ug/l  
RT: 4.472 min Scan# 435  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01




Tgt Ion: 43 Resp: 102622  
Ion Ratio Lower Upper  
43 100  
74 26.6 16.6 24.8#





#19  
Methyl tert-butyl Ether  
Concen: 50.810 ug/l  
RT: 5.216 min Scan# 511  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 73 Resp: 389740  
Ion Ratio Lower Upper  
73 100  
57 21.3 18.6 27.8

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#20  
Methylene Chloride  
Concen: 51.267 ug/l  
RT: 4.704 min Scan# 473  
Delta R.T. -0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 84 Resp: 179174  
Ion Ratio Lower Upper  
84 100  
49 111.8 114.4 171.6#  
51 35.3 35.3 52.9#  
86 62.7 53.7 80.5





#21  
trans-1,2-Dichloroethene  
Concen: 50.338 ug/l  
RT: 5.216 min Scan# 511  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 96 Resp: 16526  
Ion Ratio Lower Upper  
96 100  
61 130.6 120.5 180.7  
98 63.6 49.8 74.8

Manual Integrations  
APPROVED


Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda



11/02/2022




#22  
Diisopropyl ether  
Concen: 51.375 ug/l  
RT: 6.112 min Scan# 704  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01




Tgt Ion: 45 Resp: 465729  
Ion Ratio Lower Upper  
45 100  
43 53.5 49.3 73.9  
87 29.0 19.2 28.8#  
59 11.8 8.4 12.6





#23  
 Vinyl Acetate  
 Concen: 260.013 ug/l  
 RT: 6.051 min Scan# 694  
 Delta R.T. -0.007 min  
 Lab File: VY011217.D  
 Acq: 01 Nov 2022 13:01

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 ICVVY110122






Tgt Ion: 43 Resp: 1438678  
 Ion Ratio Lower Upper  
 43 100  
 86 12.0 7.1 10.7

Manual Integrations  
 APPROVED

Reviewed By :Krupa  
 Patel




11/02/2022  
 Supervised By :Mahesh  
 Dadoda

11/02/2022



#24  
 1,1-Dichloroethane  
 Concen: 49.809 ug/l  
 RT: 6.009 min Scan# 687  
 Delta R.T. -0.006 min  
 Lab File: VY011217.D  
 Acq: 01 Nov 2022 13:01

Tgt Ion: 63 Resp: 282772  
 Ion Ratio Lower Upper  
 63 100  
 98 7.2 3.6 10.8  
 100 4.5 2.0 5.8





#25

2-Butanone

Concen: 251.055 ug/l

RT: 6.978 min Scan# 846

Delta R.T. 0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122

Tgt Ion: 43 Resp: 302630

Ion Ratio Lower Upper

43 100

72 29.6 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :Krupa





Patel

11/02/2022

Supervised By :Mahesh

Dadoda

11/02/2022



#26

2,2-Dichloropropane

Concen: 48.695 ug/l

RT: 6.972 min Scan# 845

Delta R.T. -0.006 min

Lab File: VY011217.D




Acq: 01 Nov 2022 13:01

Tgt Ion: 77 Resp: 252778

Ion Ratio Lower Upper

77 100


97 23.5 10.4 31.4





#27  
 cis-1,2-Dichloroethene  
 Concen: 50.801 ug/l  
 RT: 6.978 min Scan# 846  
 Delta R.T. 0.000 min  
 Lab File: VY011217.D  
 Acq: 01 Nov 2022 13:01

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 ICVVY110122






Tgt Ion: 96 Resp: 183774  
 Ion Ratio Lower Upper  
 96 100  
 61 137.9 0.0 319.8  
 98 64.4 0.0 127.0

Manual Integrations  
 APPROVED


Reviewed By :Krupa  
 Patel

11/02/2022  
 Supervised By :Mahesh  
 Dadoda



11/02/2022




#28  
 Bromochloromethane  
 Concen: 50.302 ug/l  
 RT: 7.326 min Scan# 903  
 Delta R.T. -0.006 min  
 Lab File: VY011217.D  
 Acq: 01 Nov 2022 13:01



Tgt Ion: 49 Resp: 109694  
 Ion Ratio Lower Upper  
 49 100  
 129 2.1 0.0 2.2  
 130 81.1 55.4 83.0





#29

Tetrahydrofuran

Concen: 247.295 ug/l

RT: 7.344 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122

Tgt Ion: 42 Resp: 17354

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 42  | 100   |       |       |
| 72  | 47.9  | 30.3  | 45.5  |
| 71  | 45.8  | 29.0  | 43.4  |





Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel

11/02/2022

Supervised By :Mahesh  
Dadoda

11/02/2022



#30

Chloroform

Concen: 49.620 ug/l

RT: 7.502 min Scan# 932




Delta R.T. 0.000 min


Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Tgt Ion: 83 Resp: 284507


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 85  | 66.0  | 51.8  | 77.6  |





#31  
Cyclohexane  
Concen: 49.971 ug/l  
RT: 7.777 min Scan# 911  
Delta R.T. -0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 56 Resp: 26286  
Ion Ratio Lower Upper  
56 100  
69 33.6 26.7 40.1  
84 90.7 67.2 100.8

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#32  
1,1,1-Trichloroethane  
Concen: 49.511 ug/l  
RT: 7.697 min Scan# 964  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion: 97 Resp: 251796  
Ion Ratio Lower Upper  
97 100  
99 65.1 50.7 76.1  
61 40.4 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 46.654 ug/l

RT: 8.136 min Scan# 1036

Delta R.T. 0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122

Tgt Ion: 65 Resp: 132410

Ion Ratio Lower Upper

65 100

67 55.3 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Krupa





Patel

11/02/2022

Supervised By :Mahesh

Dadoda

11/02/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01




Tgt Ion:114 Resp: 424052

Ion Ratio Lower Upper

114 100

63 19.2 0.0 43.2


88 15.7 0.0 31.8





#35  
Dibromofluoromethane  
Concen: 49.064 ug/l  
RT: 7.716 min Scan# 900  
Delta R.T. 0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:113 Resp: 13367  
Ion Ratio Lower Upper  
113 100  
111 102.5 82.6 124.0  
192 19.8 17.6 26.4

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda

11/02/2022



#36  
1,1-Dichloropropene  
Concen: 51.199 ug/l  
RT: 7.911 min Scan# 999  
Delta R.T. -0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion: 75 Resp: 227798  
Ion Ratio Lower Upper  
75 100  
110 35.9 17.1 51.2  
77 31.4 24.4 36.6





#37

Ethyl Acetate

Concen: 49.665 ug/l

RT: 7.069 min Scan# 861

Delta R.T. -0.001 min

Lab File: VY011217.D


Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122



Tgt Ion: 43 Resp: 119120

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 14.9  | 9.7   | 14.5  |
| 70  | 11.3  | 7.1   | 10.7  |



Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel

11/02/2022

Supervised By :Mahesh  
Dadoda

11/02/2022



#38

Carbon Tetrachloride

Concen: 50.484 ug/l

RT: 7.899 min Scan# 997




Delta R.T. -0.000 min


Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Tgt Ion:117 Resp: 233031


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 119 | 95.5  | 77.9  | 116.9 |
| 121 | 30.0  | 24.8  | 37.2  |





#39  
Methylcyclohexane  
Concen: 53.325 ug/l  
RT: 9.179 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 83 Resp: 288644  
Ion Ratio Lower Upper  
83 100  
55 71.9 66.5 99.7  
98 48.6 37.9 56.9

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#40  
Benzene  
Concen: 50.394 ug/l  
RT: 8.155 min Scan# 1039  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 78 Resp: 658976  
Ion Ratio Lower Upper  
78 100  
77 23.5 20.4 30.6





#41  
Methacrylonitrile  
Concen: 53.484 ug/l m  
RT: 7.301 min Scan# 899  
Delta R.T. -0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 41 Resp: 6970  
Ion Ratio Lower Upper  
41 100  
39 0.0 92.9 139.3  
67 0.0 0.0 0.0  
52 0.0 0.0 0.0

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#42  
1,2-Dichloroethane  
Concen: 49.514 ug/l  
RT: 8.234 min Scan# 1052  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 62 Resp: 175820  
Ion Ratio Lower Upper  
62 100  
98 10.7 0.0 14.8





#43  
Isopropyl Acetate  
Concen: 50.141 ug/l  
RT: 8.270 min Scan# 1058  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument : MSVOA\_Y  
ClientSampleId : ICVVY110122






Tgt Ion: 43 Resp: 21188  
Ion Ratio Lower Upper  
43 100  
61 31.7 20.2 30.2  
87 15.4 8.9 13.3

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#44  
Trichloroethene  
Concen: 50.242 ug/l  
RT: 8.935 min Scan# 1167  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion:130 Resp: 175822  
Ion Ratio Lower Upper  
130 100  
95 99.0 0.0 188.2





#45

1,2-Dichloropropane

Concen: 50.010 ug/l

RT: 9.209 min Scan# 1212

Delta R.T. -0.007 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122

Tgt Ion: 63 Resp: 16167

Ion Ratio Lower Upper

63 100

65 31.3 23.0 34.6

Manual Integrations





APPROVED

Reviewed By :Krupa  
Patel

11/02/2022

Supervised By :Mahesh  
Dadoda

11/02/2022



#46

Dibromomethane

Concen: 49.773 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01




Tgt Ion: 93 Resp: 90915


Ion Ratio Lower Upper

93 100

95 84.2 64.3 96.5


174 100.6 83.4 125.0





#47  
Bromodichloromethane  
Concen: 50.381 ug/l  
RT: 9.496 min Scan# 1259  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument : MSVOA\_Y  
ClientSampleId : ICVVY110122






Tgt Ion: 83 Resp: 218150  
Ion Ratio Lower Upper  
83 100  
85 64.4 51.9 77.9  
127 9.0 6.6 10.0

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#48  
Methyl methacrylate  
Concen: 51.613 ug/l  
RT: 9.289 min Scan# 1225  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 41 Resp: 100180  
Ion Ratio Lower Upper  
41 100  
69 96.1 58.2 87.2#  
39 48.5 47.4 71.0





#49  
1,4-Dioxane  
Concen: 1065.519 ug/l  
RT: 9.295 min Scan# 1226  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 88 Resp: 2618  
Ion Ratio Lower Upper  
88 100  
43 25.4 32.8 49.2  
58 62.3 59.8 89.6

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#50  
Toluene-d8  
Concen: 48.444 ug/l  
RT: 10.179 min Scan# 1371  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 98 Resp: 512452  
Ion Ratio Lower Upper  
98 100  
100 65.2 50.2 75.2





#51  
4-Methyl-2-Pentanone  
Concen: 251.963 ug/l  
RT: 10.069 min Scan# 1353  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 43 Resp: 578929  
Ion Ratio Lower Upper  
43 100  
58 43.1 28.1 42.1

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#52  
Toluene  
Concen: 51.268 ug/l  
RT: 10.240 min Scan# 1381  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 92 Resp: 415691  
Ion Ratio Lower Upper  
92 100  
91 170.9 142.6 213.8





#53  
t-1,3-Dichloropropene  
Concen: 50.765 ug/l  
RT: 10.465 min Scan# 1418  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 75 Resp: 231495  
Ion Ratio Lower Upper  
75 100  
77 32.3 25.4 38.2

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#54  
cis-1,3-Dichloropropene  
Concen: 51.097 ug/l  
RT: 9.923 min Scan# 1329  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 75 Resp: 266619  
Ion Ratio Lower Upper  
75 100  
77 31.3 25.6 38.4  
39 38.6 47.6 71.4#





#55  
1,1,2-Trichloroethane  
Concen: 49.280 ug/l  
RT: 10.642 min Scan# 1447  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 97 Resp: 129010  
Ion Ratio Lower Upper  
97 100  
83 86.1 69.8 104.8  
85 56.8 44.3 66.5  
99 63.4 49.8 74.6

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#56  
Ethyl methacrylate  
Concen: 52.749 ug/l  
RT: 10.508 min Scan# 1425  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 69 Resp: 182405  
Ion Ratio Lower Upper  
69 100  
41 58.1 63.4 95.0#  
39 28.4 35.8 53.8#





#57  
1,3-Dichloropropane  
Concen: 49.639 ug/l  
RT: 10.788 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 76 Resp: 222432  
Ion Ratio Lower Upper  
76 100  
78 32.8 25.0 37.6

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#58  
2-Chloroethyl Vinyl ether  
Concen: 256.018 ug/l  
RT: 9.782 min Scan# 1306  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 63 Resp: 415036  
Ion Ratio Lower Upper  
63 100  
106 28.9 21.5 32.3





#59  
2-Hexanone  
Concen: 262.555 ug/l  
RT: 10.831 min Scan# 1478  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 43 Resp: 430688  
Ion Ratio Lower Upper  
43 100  
58 60.2 24.4 73.4

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#60  
Dibromochloromethane  
Concen: 49.629 ug/l  
RT: 10.983 min Scan# 1503  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:129 Resp: 154367  
Ion Ratio Lower Upper  
129 100  
127 77.3 38.7 116.1





#61  
1,2-Dibromoethane  
Concen: 49.662 ug/l  
RT: 11.087 min Scan# 1520  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122





Tgt Ion:107 Resp: 123429  
Ion Ratio Lower Upper  
107 100  
109 94.6 75.4 113.2

Manual Integrations  
APPROVED


Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda



11/02/2022




#62  
4-Bromofluorobenzene  
Concen: 49.101 ug/l  
RT: 12.477 min Scan# 1748  
Delta R.T. -0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01




Tgt Ion: 95 Resp: 180314  
Ion Ratio Lower Upper  
95 100  
174 81.5 0.0 177.0  
176 78.6 0.0 175.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.489 min Scan# 117  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:117 Resp: 38496  
Ion Ratio Lower Upper  
117 100  
82 53.8 43.0 64.4  
119 32.0 25.4 38.0

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#64  
Tetrachloroethene  
Concen: 50.838 ug/l  
RT: 10.715 min Scan# 1459  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:164 Resp: 166244  
Ion Ratio Lower Upper  
164 100  
166 124.1 100.4 150.6  
129 94.4 71.4 107.0  
131 90.5 70.8 106.2





#65  
Chlorobenzene  
Concen: 51.277 ug/l  
RT: 11.514 min Scan# 1590  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:112 Resp: 435508  
Ion Ratio Lower Upper  
112 100  
114 31.9 25.8 38.6

Manual Integrations  
APPROVED


Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda



11/02/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 50.967 ug/l  
RT: 11.587 min Scan# 1602  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01



Tgt Ion:131 Resp: 159386  
Ion Ratio Lower Upper  
131 100  
133 95.4 47.9 143.6  
119 65.4 30.8 92.3





#67  
Ethyl Benzene  
Concen: 52.881 ug/l  
RT: 11.593 min Scan# 1603  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122





Tgt Ion: 91 Resp: 796692  
Ion Ratio Lower Upper  
91 100  
106 31.3 24.5 36.7

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#68  
m/p-Xylenes  
Concen: 105.232 ug/l  
RT: 11.703 min Scan# 1621  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion:106 Resp: 615421  
Ion Ratio Lower Upper  
106 100  
91 199.5 166.6 249.8





#69  
o-Xylene  
Concen: 52.820 ug/l  
RT: 12.026 min Scan# 1674  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122





Tgt Ion:106 Resp: 28956  
Ion Ratio Lower Upper  
106 100  
91 210.8 110.3 331.0

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#70  
Styrene  
Concen: 52.675 ug/l  
RT: 12.044 min Scan# 1677  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:104 Resp: 489531  
Ion Ratio Lower Upper  
104 100  
78 47.9 42.4 63.6  
103 54.9 44.4 66.6





#71  
Bromoform  
Concen: 50.935 ug/l  
RT: 12.209 min Scan# 1704  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:173 Resp: 9900  
Ion Ratio Lower Upper  
173 100  
175 48.1 24.6 74.0  
254 0.1 0.0 0.0

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.422 min Scan# 1903  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:152 Resp: 186299  
Ion Ratio Lower Upper  
152 100  
115 58.8 28.2 84.7  
150 175.0 0.0 347.6





#73  
Isopropylbenzene  
Concen: 54.585 ug/l  
RT: 12.331 min Scan# 1724  
Delta R.T. 0.006 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 105 Resp: 760300  
Ion Ratio Lower Upper  
105 100  
120 26.7 13.4 40.2

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#74  
N-amy1 acetate  
Concen: 53.444 ug/l  
RT: 12.142 min Scan# 1693  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 43 Resp: 186864  
Ion Ratio Lower Upper  
43 100  
70 48.4 28.7 43.1#  
55 26.9 17.3 25.9#  
61 27.4 17.0 25.6#





#75  
1,1,2,2-Tetrachloroethane  
Concen: 51.159 ug/l  
RT: 12.581 min Scan# 1765  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 83 Resp: 147458  
Ion Ratio Lower Upper  
83 100  
131 10.5 5.6 16.8  
85 65.0 32.3 96.8

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#76  
1,2,3-Trichloropropane  
Concen: 48.452 ug/l m  
RT: 12.629 min Scan# 1773  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion: 75 Resp: 109891  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 53.101 ug/l

RT: 12.605 min Scan# 1769

Delta R.T. 0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 156   | Resp: | 173980 |
| Ion Ratio | Lower | Upper |        |
| 156       | 100   |       |        |
| 77        | 186.9 | 95.5  | 286.4  |
| 158       | 97.2  | 48.0  | 144.0  |





Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel

11/02/2022

Supervised By :Mahesh  
Dadoda

11/02/2022



#78

n-propylbenzene

Concen: 54.457 ug/l




RT: 12.672 min Scan# 1780


Delta R.T. 0.000 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 91    | Resp: | 928169 |
| Ion Ratio | Lower | Upper |        |
| 91        | 100   |       |        |
| 120       | 23.2  | 11.3  | 33.9   |





#79  
2-Chlorotoluene  
Concen: 53.187 ug/l  
RT: 12.757 min Scan# 1794  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 91 Resp: 519410  
Ion Ratio Lower Upper  
91 100  
126 34.2 16.9 50.6

Manual Integrations  
APPROVED


Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda



11/02/2022




#80  
1,3,5-Trimethylbenzene  
Concen: 54.470 ug/l  
RT: 12.812 min Scan# 1803  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01




Tgt Ion:105 Resp: 642019  
Ion Ratio Lower Upper  
105 100  
120 49.1 24.4 73.4





#81  
trans-1,4-Dichloro-2-butene  
Concen: 51.465 ug/l  
RT: 12.380 min Scan# 1732  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 75 Resp: 5454  
Ion Ratio Lower Upper  
75 100  
53 81.8 85.3 127.9  
89 45.3 37.0 55.4

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#82  
4-Chlorotoluene  
Concen: 53.161 ug/l  
RT: 12.855 min Scan# 1810  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 91 Resp: 535006  
Ion Ratio Lower Upper  
91 100  
126 34.1 17.0 51.0





#83  
tert-Butylbenzene  
Concen: 54.382 ug/l  
RT: 13.075 min Scan# 1846  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument : MSVOA\_Y  
ClientSampleId : ICVVY110122






Tgt Ion:119 Resp: 552304  
Ion Ratio Lower Upper  
119 100  
91 63.2 34.8 104.4  
134 26.0 13.3 39.9

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#84  
1,2,4-Trimethylbenzene  
Concen: 54.374 ug/l  
RT: 13.117 min Scan# 1853  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:105 Resp: 633862  
Ion Ratio Lower Upper  
105 100  
120 44.7 22.3 66.9





#85  
 sec-Butylbenzene  
 Concen: 54.578 ug/l  
 RT: 13.251 min Scan# 1875  
 Delta R.T. 0.000 min  
 Lab File: VY011217.D  
 Acq: 01 Nov 2022 13:01

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 ICVVY110122






Tgt Ion:105 Resp: 821835  
 Ion Ratio Lower Upper  
 105 100  
 134 20.4 10.2 30.4

Manual Integrations  
 APPROVED

Reviewed By :Krupa  
 Patel




11/02/2022  
 Supervised By :Mahesh  
 Dadoda


11/02/2022



#86  
 p-Isopropyltoluene  
 Concen: 54.859 ug/l  
 RT: 13.367 min Scan# 1894  
 Delta R.T. 0.000 min  
 Lab File: VY011217.D  
 Acq: 01 Nov 2022 13:01

Tgt Ion:119 Resp: 688525  
 Ion Ratio Lower Upper  
 119 100  
 134 26.2 13.2 39.5  
 91 22.6 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 52.051 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. 0.000 min

Lab File: VY011217.D


Acq: 01 Nov 2022 13:01

Instrument :

MSVOA\_Y

ClientSampleId :

ICVVY110122



|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 146   | Resp: | 34196 |
| Ion Ratio | Lower | Upper |       |
| 146       | 100   |       |       |
| 111       | 40.0  | 19.5  | 58.5  |
| 148       | 63.2  | 31.6  | 95.0  |




Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel

11/02/2022

Supervised By :Mahesh  
Dadoda

11/02/2022



#88

1,4-Dichlorobenzene

Concen: 51.052 ug/l




RT: 13.446 min Scan# 1907


Delta R.T. -0.001 min

Lab File: VY011217.D

Acq: 01 Nov 2022 13:01


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 146   | Resp: | 336381 |
| Ion Ratio | Lower | Upper |        |
| 146       | 100   |       |        |
| 111       | 39.2  | 19.3  | 57.8   |
| 148       | 63.7  | 32.6  | 97.8   |





#89  
n-Butylbenzene  
Concen: 54.745 ug/l  
RT: 13.696 min Scan# 1948  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion: 91 Resp: 64161  
Ion Ratio Lower Upper  
91 100  
92 53.7 26.3 78.9  
134 26.3 13.0 38.9

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#90  
Hexachloroethane  
Concen: 51.736 ug/l  
RT: 13.958 min Scan# 1991  
Delta R.T. -0.001 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:117 Resp: 129861  
Ion Ratio Lower Upper  
117 100  
201 76.9 43.0 128.8





#91  
1,2-Dichlorobenzene  
Concen: 51.864 ug/l  
RT: 13.739 min Scan# 1955  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:146 Resp: 30435  
Ion Ratio Lower Upper  
146 100  
111 41.1 20.3 60.9  
148 63.6 32.3 96.9

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 49.453 ug/l  
RT: 14.355 min Scan# 2056  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion: 75 Resp: 24418  
Ion Ratio Lower Upper  
75 100  
155 89.1 38.0 113.9  
157 113.9 47.8 143.3





#93  
1,2,4-Trichlorobenzene  
Concen: 51.894 ug/l  
RT: 15.007 min Scan# 2163  
Delta R.T. 0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:180 Resp: 184060  
Ion Ratio Lower Upper  
180 100  
182 95.1 47.3 141.9  
145 30.7 14.6 43.8

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel




11/02/2022  
Supervised By :Mahesh  
Dadoda


11/02/2022



#94  
Hexachlorobutadiene  
Concen: 51.687 ug/l  
RT: 15.111 min Scan# 2180  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01


Tgt Ion:225 Resp: 106745  
Ion Ratio Lower Upper  
225 100  
223 63.2 30.8 92.3  
227 63.1 31.6 94.7





#95  
Naphthalene  
Concen: 46.312 ug/l  
RT: 15.233 min Scan# 2200  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Instrument :  
MSVOA\_Y  
ClientSampleId :  
ICVVY110122






Tgt Ion:128 Resp: 381439  
Ion Ratio Lower Upper  
128 100  
127 13.0 11.0 16.4  
129 10.8 8.4 12.6

Manual Integrations  
APPROVED

Reviewed By :Krupa  
Patel

11/02/2022  
Supervised By :Mahesh  
Dadoda

11/02/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 50.784 ug/l  
RT: 15.422 min Scan# 2231  
Delta R.T. -0.000 min  
Lab File: VY011217.D  
Acq: 01 Nov 2022 13:01

Tgt Ion:180 Resp: 158295  
Ion Ratio Lower Upper  
180 100  
182 94.5 47.3 141.9  
145 33.4 15.3 45.9

