

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111022\
 Data File : VY011387.D
 Acq On : 10 Nov 2022 14:00
 Operator : KP/MD
 Sample : VY1110SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1110SBL01

Quant Time: Nov 10 17:51:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	282029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	465587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	407404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	190757	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	116967	49.493	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.980%
35) Dibromofluoromethane	7.716	113	124927	47.878	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.760%
50) Toluene-d8	10.179	98	440063	44.147	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.300%
62) 4-Bromofluorobenzene	12.477	95	153706	44.209	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111022\
Data File : VY011387.D
Acq On : 10 Nov 2022 14:00
Operator : KP/MD
Sample : VY1110SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1110SBL01

Quant Time: Nov 10 17:51:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

