

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111022\
 Data File : VY011398.D
 Acq On : 10 Nov 2022 18:09
 Operator : KP/MD
 Sample : N5531-10
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-30

Quant Time: Nov 10 22:18:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	241891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	415886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	376311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	161833	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	175713	86.687	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	173.380%#
35) Dibromofluoromethane	7.716	113	176040	75.530	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	151.060%#
50) Toluene-d8	10.179	98	650953	73.108	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	146.220%#
62) 4-Bromofluorobenzene	12.477	95	201517	64.887	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	129.780%

Target Compounds			Qvalue			
16) Acetone	3.942	43	14833	27.732	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111022\
Data File : VY011398.D
Acq On : 10 Nov 2022 18:09
Operator : KP/MD
Sample : N5531-10
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-30

Quant Time: Nov 10 22:18:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

