

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111022\
 Data File : VY011401.D
 Acq On : 10 Nov 2022 19:17
 Operator : KP/MD
 Sample : N5531-15
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-27

Quant Time: Nov 10 22:18:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	24163	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	40349	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	34598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	11533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	23433	115.731	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 231.460%#	
35) Dibromofluoromethane	7.716	113	19900	88.004	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 176.000%#	
50) Toluene-d8	10.179	98	58960	68.251	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 136.500%	
62) 4-Bromofluorobenzene	12.477	95	16113	53.477	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.960%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	4487	9.360	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111022\
Data File : VY011401.D
Acq On : 10 Nov 2022 19:17
Operator : KP/MD
Sample : N5531-15
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-27

Quant Time: Nov 10 22:18:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

