

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111522\
 Data File : VY011461.D
 Acq On : 15 Nov 2022 16:22
 Operator : KP/MD
 Sample : N5617-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-12013

Quant Time: Nov 16 00:21:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	334260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	567683	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	472424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	185567	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	172525	61.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.180%
35) Dibromofluoromethane	7.722	113	186121	58.502	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.000%
50) Toluene-d8	10.179	98	644997	53.069	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.140%
62) 4-Bromofluorobenzene	12.483	95	196605	46.378	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111522\
Data File : VY011461.D
Acq On : 15 Nov 2022 16:22
Operator : KP/MD
Sample : N5617-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-12013

Quant Time: Nov 16 00:21:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

