



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111522\
 Data File : VY011449.D
 Acq On : 15 Nov 2022 11:06
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Title : SW846 8260
 Last Update : Thu Nov 10 17:47:26 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.4	18353	PASS
75	95	30	60	47.8	56995	PASS
95	95	100	100	100.0	119213	PASS
96	95	5	9	6.5	7799	PASS
173	174	0.00	2	0.9	949	PASS
174	95	50	100	86.2	102717	PASS
175	174	5	9	7.3	7502	PASS
176	174	95	101	95.1	97672	PASS
177	176	5	9	6.8	6671	PASS