

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
 Data File : VY011507.D
 Acq On : 18 Nov 2022 17:24
 Operator : KP/MD
 Sample : N5641-20
 Misc : 6.20g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-PILE-20

Quant Time: Nov 19 22:41:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	186791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	354869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	335629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	147068	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	132486	69.852	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.700%
35) Dibromofluoromethane	7.722	113	115128	52.701	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.400%
50) Toluene-d8	10.179	98	434240	54.701	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.400%
62) 4-Bromofluorobenzene	12.483	95	137912	46.987	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
Data File : VY011507.D
Acq On : 18 Nov 2022 17:24
Operator : KP/MD
Sample : N5641-20
Misc : 6.20g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-PILE-20

Quant Time: Nov 19 22:41:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

