

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
 Data File : VY011510.D
 Acq On : 18 Nov 2022 18:34
 Operator : KP/MD
 Sample : N5696-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BANK-RUN-SOIL-SAMPLE-3

Quant Time: Nov 19 22:42:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	3632	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	7039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	7690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	4345	117.817	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 235.640%#	
35) Dibromofluoromethane	7.722	113	1857	42.856	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 85.720%	
50) Toluene-d8	10.185	98	8421	53.479	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.960%	
62) 4-Bromofluorobenzene	12.489	95	2605	44.745	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
Data File : VY011510.D
Acq On : 18 Nov 2022 18:34
Operator : KP/MD
Sample : N5696-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-3

Quant Time: Nov 19 22:42:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

