

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
 Data File : VY011511.D
 Acq On : 18 Nov 2022 18:58
 Operator : KP/MD
 Sample : N5698-02
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BANK-RUN-SOIL-SAMPLE-2

Quant Time: Nov 19 22:42:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	179696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	344073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	327017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	137397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	122478	67.125	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 134.240%	
35) Dibromofluoromethane	7.722	113	112337	53.037	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.080%	
50) Toluene-d8	10.185	98	420454	54.626	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.260%	
62) 4-Bromofluorobenzene	12.483	95	130691	45.924	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 91.840%	
Target Compounds						Qvalue
16) Acetone	3.954	43	10670	18.003	ug/l	94
17) Carbon Disulfide	4.198	76	51791	11.833	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
Data File : VY011511.D
Acq On : 18 Nov 2022 18:58
Operator : KP/MD
Sample : N5698-02
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BANK-RUN-SOIL-SAMPLE-2

Quant Time: Nov 19 22:42:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

