



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111822\
Data File : VY011490.D
Acq On : 18 Nov 2022 09:10
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Title : SW846 8260
Last Update : Sat Nov 19 22:37:14 2022



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.4	10922	PASS
75	95	30	60	52.6	28147	PASS
95	95	100	100	100.0	53485	PASS
96	95	5	9	6.4	3446	PASS
173	174	0.00	2	1.3	622	PASS
174	95	50	100	88.9	47547	PASS
175	174	5	9	7.6	3598	PASS
176	174	95	101	97.1	46171	PASS
177	176	5	9	7.0	3240	PASS