

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112822\
 Data File : VY011595.D
 Acq On : 28 Nov 2022 12:02
 Operator : KP/MD
 Sample : N5766-04RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CR-4RE

Quant Time: Nov 29 00:24:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	368	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.691	114	635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	386	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	60	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	454	121.499	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	243.000%#
35) Dibromofluoromethane	7.722	113	187	47.838	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.680%
50) Toluene-d8	10.185	98	597	42.027	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	84.060%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
3) Chloromethane	2.119	50	19	1.802	ug/l #	44
10) Methyl Iodide	4.107	142	144	35.009	ug/l #	35

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112822\
Data File : VY011595.D
Acq On : 28 Nov 2022 12:02
Operator : KP/MD
Sample : N5766-04RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CR-4RE

Quant Time: Nov 29 00:24:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

