

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112822\
 Data File : VY011605.D
 Acq On : 28 Nov 2022 17:30
 Operator : KP/MD
 Sample : N5787-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-24-STOCK

Quant Time: Nov 29 00:26:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	488	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	372	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	333	67.203	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.400%
35) Dibromofluoromethane	7.728	113	194	44.138	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.280%
50) Toluene-d8	10.179	98	616	38.567	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.140%
62) 4-Bromofluorobenzene	12.489	95	49	8.297	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	16.600%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY112822\
Data File : VY011605.D
Acq On : 28 Nov 2022 17:30
Operator : KP/MD
Sample : N5787-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-24-STOCK

Quant Time: Nov 29 00:26:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

