



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120622\
 Data File : VY011709.D
 Acq On : 06 Dec 2022 09:01
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Title : SW846 8260
 Last Update : Sat Nov 19 22:37:14 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	13141	PASS
75	95	30	60	52.3	36960	PASS
95	95	100	100	100.0	70675	PASS
96	95	5	9	6.7	4715	PASS
173	174	0.00	2	1.3	789	PASS
174	95	50	100	85.0	60069	PASS
175	174	5	9	7.1	4254	PASS
176	174	95	101	96.1	57701	PASS
177	176	5	9	7.5	4305	PASS