

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120722\
 Data File : VY011731.D
 Acq On : 07 Dec 2022 14:17
 Operator : KP/MD
 Sample : N5897-09
 Misc : 6.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STOCK-PILE

Quant Time: Dec 08 07:56:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 19 22:37:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	7662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	12058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	8939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3483	44.769	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.540%
35) Dibromofluoromethane	7.709	113	2086	28.103	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	56.200%
50) Toluene-d8	10.178	98	13461	49.904	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.800%
62) 4-Bromofluorobenzene	12.483	95	2817	28.246	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	56.500%
Target Compounds						Qvalue
16) Acetone	3.942	43	737	29.164	ug/l	# 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120722\
Data File : VY011731.D
Acq On : 07 Dec 2022 14:17
Operator : KP/MD
Sample : N5897-09
Misc : 6.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STOCK-PILE

Quant Time: Dec 08 07:56:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111822S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 19 22:37:14 2022
Response via : Initial Calibration

