

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
 Data File : VY011744.D
 Acq On : 09 Dec 2022 15:35
 Operator : KP/MD
 Sample : VY1209SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1209SBL01

Quant Time: Dec 10 03:36:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 03:30:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	244101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	400870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	347346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	160011	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	85659	51.576	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.160%
35) Dibromofluoromethane	7.716	113	87274	48.917	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.840%
50) Toluene-d8	10.179	98	265856	44.844	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.680%
62) 4-Bromofluorobenzene	12.483	95	103666	44.061	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
Data File : VY011744.D
Acq On : 09 Dec 2022 15:35
Operator : KP/MD
Sample : VY1209SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1209SBL01

Quant Time: Dec 10 03:36:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 03:30:20 2022
Response via : Initial Calibration

