

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
 Data File : VY011750.D
 Acq On : 09 Dec 2022 18:03
 Operator : KP/MD
 Sample : N5943-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RM-301

Quant Time: Dec 10 03:56:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 03:30:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	3556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	6082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	6241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1334	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3024	132.532	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	265.060%#
35) Dibromofluoromethane	7.722	113	929	33.062	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	66.120%
50) Toluene-d8	10.179	98	7410	88.790	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	177.580%#
62) 4-Bromofluorobenzene	12.483	95	1968	56.395	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
Data File : VY011750.D
Acq On : 09 Dec 2022 18:03
Operator : KP/MD
Sample : N5943-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RM-301

Quant Time: Dec 10 03:56:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 03:30:20 2022
Response via : Initial Calibration

