

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
 Data File : VY011753.D
 Acq On : 09 Dec 2022 19:14
 Operator : KP/MD
 Sample : N5963-04
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR200046-03

Quant Time: Dec 10 03:57:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 03:30:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	47184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	78531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	69543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	25734	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	26617	86.131	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	172.260%#
35) Dibromofluoromethane	7.716	113	26607	78.472	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	156.940%#
50) Toluene-d8	10.179	98	91723	84.803	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	169.600%#
62) 4-Bromofluorobenzene	12.483	95	24317	53.751	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
Data File : VY011753.D
Acq On : 09 Dec 2022 19:14
Operator : KP/MD
Sample : N5963-04
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR200046-03

Quant Time: Dec 10 03:57:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 03:30:20 2022
Response via : Initial Calibration

