



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY120922\
Data File : VY011735.D
Acq On : 09 Dec 2022 09:59
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Title : SW846 8260
Last Update : Sat Dec 10 04:19:59 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.3	17589	PASS
75	95	30	60	52.6	50613	PASS
95	95	100	100	100.0	96229	PASS
96	95	5	9	6.8	6525	PASS
173	174	0.00	2	1.1	888	PASS
174	95	50	100	85.1	81923	PASS
175	174	5	9	7.1	5787	PASS
176	174	95	101	95.5	78203	PASS
177	176	5	9	6.9	5385	PASS