

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
 Data File : VY011764.D
 Acq On : 12 Dec 2022 12:33
 Operator : KP/MD
 Sample : N5942-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-120722

Quant Time: Dec 12 13:27:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	181	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.703	114	343	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.483	117	48	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79	65.442	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.880%	
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	= 0.000%#	
50) Toluene-d8	10.179	98	20	-3.039	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= -6.080%#	
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	= 0.000%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
Data File : VY011764.D
Acq On : 12 Dec 2022 12:33
Operator : KP/MD
Sample : N5942-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-120722

Quant Time: Dec 12 13:27:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

