

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
 Data File : VY011766.D
 Acq On : 12 Dec 2022 13:18
 Operator : KP/MD
 Sample : N5963-01RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-1195-08RE

Quant Time: Dec 13 00:19:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	26660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	43797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	39292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	14796	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	13214	75.035	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.060%
35) Dibromofluoromethane	7.722	113	12392	64.837	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.680%
50) Toluene-d8	10.179	98	29666	45.964	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.920%
62) 4-Bromofluorobenzene	12.483	95	10764	41.626	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
Data File : VY011766.D
Acq On : 12 Dec 2022 13:18
Operator : KP/MD
Sample : N5963-01RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-1195-08RE

Quant Time: Dec 13 00:19:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

