

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
 Data File : VY011768.D
 Acq On : 12 Dec 2022 14:03
 Operator : KP/MD
 Sample : N5999-01RE
 Misc : 3.21g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-20RE

Quant Time: Dec 13 00:19:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	93619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	157622	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	143719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62366	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	40253	64.389	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.780%
35) Dibromofluoromethane	7.722	113	39887	57.543	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.080%
50) Toluene-d8	10.179	98	108188	46.679	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.360%
62) 4-Bromofluorobenzene	12.483	95	40501	43.748	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
Data File : VY011768.D
Acq On : 12 Dec 2022 14:03
Operator : KP/MD
Sample : N5999-01RE
Misc : 3.21g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-20RE

Quant Time: Dec 13 00:19:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

