

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
 Data File : VY011775.D
 Acq On : 12 Dec 2022 16:42
 Operator : KP/MD
 Sample : N5997-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 2803-COMP

Quant Time: Dec 13 00:20:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	271907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	456497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	395226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	173054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	95931	51.883	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.760%
35) Dibromofluoromethane	7.722	113	96856	47.566	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.140%
50) Toluene-d8	10.179	98	295390	43.568	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.140%
62) 4-Bromofluorobenzene	12.483	95	109872	40.660	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121222\
Data File : VY011775.D
Acq On : 12 Dec 2022 16:42
Operator : KP/MD
Sample : N5997-03
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
2803-COMP

Quant Time: Dec 13 00:20:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

