



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121520\
 Data File : VY003767.D
 Acq On : 15 Dec 2020 09:44
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/10ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y121520S.M
 Title : SW846 8260
 Last Update : Wed Dec 16 05:00:50 2020



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	16213	PASS
75	95	30	60	51.4	41565	PASS
95	95	100	100	100.0	80821	PASS
96	95	5	9	6.8	5532	PASS
173	174	0.00	2	1.2	883	PASS
174	95	50	100	87.4	70643	PASS
175	174	5	9	7.5	5293	PASS
176	174	95	101	96.4	68123	PASS
177	176	5	9	7.1	4818	PASS