

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011866.D
 Acq On : 15 Dec 2022 13:06
 Operator : KP/MD
 Sample : N6038-13
 Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-35-(5-7)

Quant Time: Dec 16 00:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	288682	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	475720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	414342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	184470	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	100417	51.079	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.160%	
35) Dibromofluoromethane	7.722	113	114706	54.631	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.260%	
50) Toluene-d8	10.179	98	323313	46.145	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.280%	
62) 4-Bromofluorobenzene	12.483	95	156077	57.251	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.500%	
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	39511	9.299	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011866.D
Acq On : 15 Dec 2022 13:06
Operator : KP/MD
Sample : N6038-13
Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-35-(5-7)

Quant Time: Dec 16 00:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

