

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011869.D
 Acq On : 15 Dec 2022 14:14
 Operator : KP/MD
 Sample : N6038-15
 Misc : 4.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-36-(3-5)

Quant Time: Dec 16 00:14:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	302427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	500635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	433794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	191867	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	106192	51.611	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.220%
35) Dibromofluoromethane	7.716	113	121870	55.194	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.380%
50) Toluene-d8	10.179	98	348043	47.378	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.760%
62) 4-Bromofluorobenzene	12.483	95	167148	58.349	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.700%
Target Compounds						
20) Methylene Chloride	4.716	84	47304	11.577	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011869.D
Acq On : 15 Dec 2022 14:14
Operator : KP/MD
Sample : N6038-15
Misc : 4.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-36-(3-5)

Quant Time: Dec 16 00:14:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

