

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011874.D
 Acq On : 15 Dec 2022 16:07
 Operator : KP/MD
 Sample : N6038-20
 Misc : 4.93g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-39-(0-2)

Quant Time: Dec 16 00:15:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	268823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	441647	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	330097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	106084	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	96255	52.734	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.460%
35) Dibromofluoromethane	7.722	113	109859	56.492	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.980%
50) Toluene-d8	10.179	98	290930	44.491	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.483	95	106389	40.699	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011874.D
Acq On : 15 Dec 2022 16:07
Operator : KP/MD
Sample : N6038-20
Misc : 4.93g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-39-(0-2)

Quant Time: Dec 16 00:15:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

