

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011878.D
 Acq On : 15 Dec 2022 17:38
 Operator : KP/MD
 Sample : N6038-24
 Misc : 3.50g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-41-(0-2)

Quant Time: Dec 16 00:15:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	265426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	421654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	303050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	81779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	91135	50.350	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.700%	
35) Dibromofluoromethane	7.715	113	106569	57.467	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.940%	
50) Toluene-d8	10.178	98	271752	43.363	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 86.720%	
62) 4-Bromofluorobenzene	12.483	95	88933	35.009	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 70.020%	
Target Compounds						
20) Methylene Chloride	4.716	84	33217	7.932	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011878.D
Acq On : 15 Dec 2022 17:38
Operator : KP/MD
Sample : N6038-24
Misc : 3.50g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-41-(0-2)

Quant Time: Dec 16 00:15:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

