

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011879.D
 Acq On : 15 Dec 2022 18:00
 Operator : KP/MD
 Sample : N6038-25
 Misc : 4.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-41-(3-5)

Quant Time: Dec 16 00:15:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	7641	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	12118	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	9522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3173	62.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.020%
35) Dibromofluoromethane	7.722	113	1927	34.593	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	69.180%
50) Toluene-d8	10.185	98	8070	45.062	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.120%
62) 4-Bromofluorobenzene	12.489	95	2802	38.865	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.720%
Target Compounds						
20) Methylene Chloride	4.710	84	1292	13.055	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011879.D
Acq On : 15 Dec 2022 18:00
Operator : KP/MD
Sample : N6038-25
Misc : 4.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-41-(3-5)

Quant Time: Dec 16 00:15:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

