

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011883.D
 Acq On : 15 Dec 2022 19:31
 Operator : KP/MD
 Sample : N6038-10
 Misc : 5.24g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-33-(5-7)

Quant Time: Dec 16 00:16:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	253049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	405554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	301714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	84867	49.058	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.120%
35) Dibromofluoromethane	7.716	113	98712	55.187	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.380%
50) Toluene-d8	10.179	98	255078	42.134	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	84.260%
62) 4-Bromofluorobenzene	12.483	95	88413	36.355	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.720%
Target Compounds						
20) Methylene Chloride	4.716	84	39065	11.340	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011883.D
Acq On : 15 Dec 2022 19:31
Operator : KP/MD
Sample : N6038-10
Misc : 5.24g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-33-(5-7)

Quant Time: Dec 16 00:16:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

