

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011887.D
 Acq On : 15 Dec 2022 21:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 16 00:17:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.273	0.215	21.2	84	0.00
3 P	Chloromethane	0.306	0.261	14.7	81	0.00
4 C	Vinyl Chloride	0.358	0.304	15.1#	79	0.00
5 T	Bromomethane	0.247	0.223	9.7	85	0.00
6 T	Chloroethane	0.248	0.220	11.3	81	0.00
7 T	Trichlorofluoromethane	0.687	0.630	8.3	83	0.00
8 T	Diethyl Ether	0.241	0.227	5.8	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.410	0.380	7.3	82	0.00
10 T	Methyl Iodide	0.480	0.461	4.0	80	0.00
11 T	Tert butyl alcohol	0.061	0.040	34.4#	82	0.00
12 CM	1,1-Dichloroethene	0.370	0.338	8.6#	81	0.00
13 T	Acrolein	0.070	0.054	22.9	68	0.00
14 T	Allyl chloride	0.615	0.564	8.3	80	0.00
15 T	Acrylonitrile	0.129	0.123	4.7	85	0.00
16 T	Acetone	0.137	0.101	26.3#	66	0.00
17 T	Carbon Disulfide	0.845	0.718	15.0	77	0.00
18 T	Methyl Acetate	0.386	0.289	25.1#	83	0.00
19 T	Methyl tert-butyl Ether	1.159	1.159	0.0	86	0.00
20 T	Methylene Chloride	0.626	0.546	12.8	97	0.00
21 T	trans-1,2-Dichloroethene	0.416	0.386	7.2	82	0.00
22 T	Diisopropyl ether	1.341	1.354	-1.0	84	0.00
23 T	Vinyl Acetate	0.765	0.795	-3.9	83	0.00
24 P	1,1-Dichloroethane	0.815	0.786	3.6	84	0.00
25 T	2-Butanone	0.176	0.150	14.8	77	0.00
26 T	2,2-Dichloropropane	0.767	0.720	6.1	82	0.00
27 T	cis-1,2-Dichloroethene	0.508	0.510	-0.4	87	0.00
28 T	Bromochloromethane	0.297	0.299	-0.7	87	0.00
29 T	Tetrahydrofuran	0.100	0.094	6.0	84	0.00
30 C	Chloroform	0.871	0.875	-0.5#	86	0.00
31 T	Cyclohexane	0.653	0.563	13.8	79	0.00
32 T	1,1,1-Trichloroethane	0.777	0.772	0.6	86	0.00
33 S	1,2-Dichloroethane-d4	0.406	0.333	18.0	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.252	0.226	10.3	88	0.00
36 T	1,1-Dichloropropene	0.379	0.361	4.7	83	0.00
37 T	Ethyl Acetate	0.222	0.207	6.8	82	0.00
38 T	Carbon Tetrachloride	0.446	0.445	0.2	85	0.00
39 T	Methylcyclohexane	0.427	0.398	6.8	80	0.00
40 TM	Benzene	1.134	1.112	1.9	85	0.00
41 T	Methacrylonitrile	0.112	0.113	-0.9	85	0.00
42 TM	1,2-Dichloroethane	0.351	0.351	0.0	87	0.00
43 T	Isopropyl Acetate	0.402	0.387	3.7	85	0.00
44 TM	Trichloroethene	0.317	0.306	3.5	85	0.00
45 C	1,2-Dichloropropane	0.293	0.290	1.0#	85	0.00
46 T	Dibromomethane	0.164	0.164	0.0	88	0.00
47 T	Bromodichloromethane	0.424	0.439	-3.5	89	0.00
48 T	Methyl methacrylate	0.177	0.171	3.4	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
 Data File : VY011887.D
 Acq On : 15 Dec 2022 21:01
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 16 00:17:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	86	0.00
50 S	Toluene-d8	0.885	0.677	23.5	82	0.00
51 T	4-Methyl-2-Pentanone	0.222	0.221	0.5	86	0.00
52 CM	Toluene	0.721	0.730	-1.2#	87	0.00
53 T	t-1,3-Dichloropropene	0.424	0.427	-0.7	87	0.00
54 T	cis-1,3-Dichloropropene	0.476	0.476	0.0	86	0.00
55 T	1,1,2-Trichloroethane	0.243	0.251	-3.3	91	0.00
56 T	Ethyl methacrylate	0.313	0.321	-2.6	86	0.00
57 T	1,3-Dichloropropane	0.405	0.411	-1.5	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.145	0.149	-2.8	87	0.00
59 T	2-Hexanone	0.158	0.150	5.1	81	0.00
60 T	Dibromochloromethane	0.304	0.319	-4.9	92	0.00
61 T	1,2-Dibromoethane	0.220	0.221	-0.5	88	0.00
62 S	4-Bromofluorobenzene	0.340	0.295	13.2	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.354	0.348	1.7	88	0.00
65 PM	Chlorobenzene	0.901	0.888	1.4	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.350	0.362	-3.4	90	0.00
67 C	Ethyl Benzene	1.587	1.583	0.3#	86	0.00
68 T	m/p-Xylenes	0.608	0.610	-0.3	87	0.00
69 T	o-Xylene	0.579	0.594	-2.6	89	0.00
70 T	Styrene	0.989	1.018	-2.9	88	0.00
71 P	Bromoform	0.216	0.224	-3.7	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.240	3.283	-1.3	88	0.00
74 T	N-amyl acetate	0.779	0.748	4.0	85	0.00
75 P	1,1,2,2-Tetrachloroethane	0.631	0.621	1.6	90	0.00
76 T	1,2,3-Trichloropropane	0.506	0.490	3.2	88	0.00
77 T	Bromobenzene	0.756	0.771	-2.0	90	0.00
78 T	n-propylbenzene	3.882	3.892	-0.3	88	0.00
79 T	2-Chlorotoluene	2.216	2.213	0.1	88	0.00
80 T	1,3,5-Trimethylbenzene	2.738	2.779	-1.5	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.223	0.211	5.4	84	0.00
82 T	4-Chlorotoluene	2.298	2.301	-0.1	89	0.00
83 T	tert-Butylbenzene	2.384	2.471	-3.6	89	0.00
84 T	1,2,4-Trimethylbenzene	2.696	2.757	-2.3	90	0.00
85 T	sec-Butylbenzene	3.551	3.604	-1.5	88	0.00
86 T	p-Isopropyltoluene	2.955	3.035	-2.7	89	0.00
87 T	1,3-Dichlorobenzene	1.522	1.538	-1.1	91	0.00
88 T	1,4-Dichlorobenzene	1.513	1.510	0.2	91	0.00
89 T	n-Butylbenzene	2.687	2.702	-0.6	87	0.00
90 T	Hexachloroethane	0.587	0.592	-0.9	92	0.00
91 T	1,2-Dichlorobenzene	1.370	1.383	-0.9	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.110	0.105	4.5	89	0.00
93 T	1,2,4-Trichlorobenzene	0.799	0.830	-3.9	93	0.00
94 T	Hexachlorobutadiene	0.508	0.513	-1.0	91	0.00
95 T	Naphthalene	1.547	1.664	-7.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121522\
Data File : VY011887.D
Acq On : 15 Dec 2022 21:01
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Dec 16 00:17:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.690	0.723	-4.8	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6