

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
 Data File : VY007157.D
 Acq On : 14 Jan 2022 10:28
 Operator : SY/MD
 Sample : VY0114SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0114SBL01

Quant Time: Jan 14 13:45:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	108926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	201511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	188911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83131	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	69854	52.265	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.520%
35) Dibromofluoromethane	7.722	113	66717	48.963	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.920%
50) Toluene-d8	10.179	98	247222	49.655	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.483	95	91573	51.706	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
Data File : VY007157.D
Acq On : 14 Jan 2022 10:28
Operator : SY/MD
Sample : VY0114SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0114SBL01

Quant Time: Jan 14 13:45:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

