

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
 Data File : VY007161.D
 Acq On : 14 Jan 2022 12:14
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 14 14:38:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	110671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	203143	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	193658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	85502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	70840	52.167	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.340%	
35) Dibromofluoromethane	7.716	113	67353	49.033	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.060%	
50) Toluene-d8	10.179	98	250285	49.866	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.740%	
62) 4-Bromofluorobenzene	12.477	95	92813	51.985	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.960%	
Target Compounds						
20) Methylene Chloride	4.710	84	3966	2.721	ug/l	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
Data File : VY007161.D
Acq On : 14 Jan 2022 12:14
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 14 14:38:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

