

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
 Data File : VY007164.D
 Acq On : 14 Jan 2022 13:23
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 14 14:39:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	106976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	199843	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	190708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83376	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	69837	53.205	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.400%
35) Dibromofluoromethane	7.715	113	66567	49.261	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.520%
50) Toluene-d8	10.178	98	244321	49.482	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.960%
62) 4-Bromofluorobenzene	12.483	95	91917	52.333	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
Data File : VY007164.D
Acq On : 14 Jan 2022 13:23
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 14 14:39:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

