

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
 Data File : VY007166.D
 Acq On : 14 Jan 2022 14:10
 Operator : SY/MD
 Sample : N1099-09
 Misc : 6.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-09

Quant Time: Jan 17 04:52:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	108116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	203617	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	193500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	85533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	72444	54.609	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.220%	
35) Dibromofluoromethane	7.716	113	68655	49.865	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.720%	
50) Toluene-d8	10.179	98	247582	49.213	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.420%	
62) 4-Bromofluorobenzene	12.477	95	94310	52.700	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.400%	
Target Compounds						Qvalue
16) Acetone	3.942	43	8184	23.771	ug/l	93
30) Chloroform	7.509	83	13335	5.309	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011422\
Data File : VY007166.D
Acq On : 14 Jan 2022 14:10
Operator : SY/MD
Sample : N1099-09
Misc : 6.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VACLOT-BACKFILL-TILCON-09

Quant Time: Jan 17 04:52:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

