

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
 Data File : VY007217.D
 Acq On : 19 Jan 2022 12:03
 Operator : SY/MD
 Sample : VY0119SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0119SBL01

Quant Time: Jan 20 00:15:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	99384	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	185603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	171695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	73284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	64047	52.521	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.040%
35) Dibromofluoromethane	7.722	113	62343	49.675	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.340%
50) Toluene-d8	10.185	98	227376	49.583	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.160%
62) 4-Bromofluorobenzene	12.483	95	81374	49.885	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.780%
Target Compounds						Qvalue
30) Chloroform	7.509	83	3048	1.320	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
Data File : VY007217.D
Acq On : 19 Jan 2022 12:03
Operator : SY/MD
Sample : VY0119SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0119SBL01

Quant Time: Jan 20 00:15:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

