

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
 Data File : VY007221.D
 Acq On : 19 Jan 2022 14:05
 Operator : SY/MD
 Sample : N1158-02
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-1A

Quant Time: Jan 20 00:16:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	87	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	8.703	114	218	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	152	50.000	ug/l	# 0.02
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	122	114.286	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 228.580%#	
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	= 0.000%#	
50) Toluene-d8	10.185	98	280	51.985	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.960%	
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	= 0.000%#	
Target Compounds						
					Qvalue	
3) Chloromethane	2.131	50	48	75.032	ug/l	# 41
67) Ethyl Benzene	11.587	91	20	3.202	ug/l	# 45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
Data File : VY007221.D
Acq On : 19 Jan 2022 14:05
Operator : SY/MD
Sample : N1158-02
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-1A

Quant Time: Jan 20 00:16:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

