

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
 Data File : VY007223.D
 Acq On : 19 Jan 2022 15:20
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 20 00:17:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	101703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	189491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	176724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	62814	50.335	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.680%	
35) Dibromofluoromethane	7.722	113	62705	48.938	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.880%	
50) Toluene-d8	10.179	98	232159	49.587	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.180%	
62) 4-Bromofluorobenzene	12.483	95	82232	49.377	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.760%	
Target Compounds						
30) Chloroform	7.514	83	4806	2.034	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
Data File : VY007223.D
Acq On : 19 Jan 2022 15:20
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 20 00:17:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

