

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
 Data File : VY007224.D
 Acq On : 19 Jan 2022 16:35
 Operator : SY/MD
 Sample : N1189-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MTR-SOIL-1

Quant Time: Jan 20 00:17:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	90746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	170700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	164539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	64399	57.837	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.680%
35) Dibromofluoromethane	7.716	113	58348	50.551	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.100%
50) Toluene-d8	10.179	98	209045	49.565	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.140%
62) 4-Bromofluorobenzene	12.483	95	77898	51.923	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.840%
Target Compounds						
16) Acetone	3.954	43	2229	7.714	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
Data File : VY007224.D
Acq On : 19 Jan 2022 16:35
Operator : SY/MD
Sample : N1189-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MTR-SOIL-1

Quant Time: Jan 20 00:17:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

