

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
 Data File : VY007216.D
 Acq On : 19 Jan 2022 10:49
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 00:15:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.175	0.142	18.9	69	0.00
3 P	Chloromethane	0.368	0.343	6.8	88	0.00
4 C	Vinyl Chloride	0.535	0.509	4.9#	81	0.00
5 T	Bromomethane	0.346	0.367	-6.1	86	0.00
6 T	Chloroethane	0.334	0.334	0.0	84	0.00
7 T	Trichlorofluoromethane	0.577	0.570	1.2	85	0.00
8 T	Diethyl Ether	0.318	0.319	-0.3	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.562	0.539	4.1	82	0.00
10 T	Methyl Iodide	0.831	0.794	4.5	81	0.00
11 T	Tert butyl alcohol	0.058	0.052	10.3	89	0.00
12 CM	1,1-Dichloroethene	0.566	0.529	6.5#	82	0.00
13 T	Acrolein	0.015	0.011	26.7#	86	0.00
14 T	Allyl chloride	0.907	0.903	0.4	85	0.00
15 T	Acrylonitrile	0.153	0.161	-5.2	88	0.00
16 T	Acetone	0.159	0.146	8.2	78	0.00
17 T	Carbon Disulfide	1.571	1.459	7.1	81	0.00
18 T	Methyl Acetate	0.567	0.554	2.3	89	0.00
19 T	Methyl tert-butyl Ether	1.114	1.190	-6.8	91	0.00
20 T	Methylene Chloride	0.658	0.588	10.6	83	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.570	7.0	81	0.00
22 T	Diisopropyl ether	1.838	1.883	-2.4	86	0.00
23 T	Vinyl Acetate	1.081	1.166	-7.9	88	0.00
24 P	1,1-Dichloroethane	1.118	1.091	2.4	83	0.00
25 T	2-Butanone	0.206	0.206	0.0	85	0.00
26 T	2,2-Dichloropropane	0.815	0.803	1.5	86	0.00
27 T	cis-1,2-Dichloroethene	0.696	0.669	3.9	82	0.00
28 T	Bromochloromethane	0.404	0.447	-10.6	85	0.00
29 T	Tetrahydrofuran	0.132	0.139	-5.3	89	0.00
30 C	Chloroform	1.162	1.151	0.9#	84	0.00
31 T	Cyclohexane	1.037	0.937	9.6	82	0.00
32 T	1,1,1-Trichloroethane	0.993	0.980	1.3	84	0.00
33 S	1,2-Dichloroethane-d4	0.614	0.611	0.5	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.338	0.325	3.8	85	0.00
36 T	1,1-Dichloropropene	0.540	0.507	6.1	83	0.00
37 T	Ethyl Acetate	0.277	0.284	-2.5	87	0.00
38 T	Carbon Tetrachloride	0.568	0.537	5.5	83	0.00
39 T	Methylcyclohexane	0.659	0.603	8.5	80	0.00
40 TM	Benzene	1.440	1.360	5.6	82	0.00
41 T	Methacrylonitrile	0.160	0.155	3.1	94	0.01
42 TM	1,2-Dichloroethane	0.454	0.457	-0.7	87	0.00
43 T	Isopropyl Acetate	0.539	0.567	-5.2	90	0.00
44 TM	Trichloroethene	0.390	0.367	5.9	81	0.00
45 C	1,2-Dichloropropane	0.366	0.357	2.5#	85	0.00
46 T	Dibromomethane	0.217	0.212	2.3	85	0.00
47 T	Bromodichloromethane	0.532	0.536	-0.8	86	0.00
48 T	Methyl methacrylate	0.245	0.248	-1.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
 Data File : VY007216.D
 Acq On : 19 Jan 2022 10:49
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 00:15:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	87	0.01
50 S	Toluene-d8	1.235	1.184	4.1	85	0.00
51 T	4-Methyl-2-Pentanone	0.273	0.286	-4.8	89	0.00
52 CM	Toluene	0.913	0.887	2.8#	83	0.00
53 T	t-1,3-Dichloropropene	0.562	0.563	-0.2	84	0.00
54 T	cis-1,3-Dichloropropene	0.629	0.617	1.9	83	0.00
55 T	1,1,2-Trichloroethane	0.294	0.284	3.4	83	0.00
56 T	Ethyl methacrylate	0.414	0.422	-1.9	87	0.00
57 T	1,3-Dichloropropane	0.512	0.510	0.4	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.188	0.180	4.3	85	0.00
59 T	2-Hexanone	0.186	0.189	-1.6	89	0.00
60 T	Dibromochloromethane	0.361	0.358	0.8	85	0.00
61 T	1,2-Dibromoethane	0.289	0.284	1.7	84	0.00
62 S	4-Bromofluorobenzene	0.439	0.419	4.6	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.360	0.330	8.3	81	0.00
65 PM	Chlorobenzene	1.093	1.055	3.5	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.405	0.397	2.0	82	0.00
67 C	Ethyl Benzene	2.055	1.984	3.5#	83	0.00
68 T	m/p-Xylenes	0.764	0.734	3.9	82	0.00
69 T	o-Xylene	0.719	0.695	3.3	83	0.00
70 T	Styrene	1.213	1.205	0.7	84	0.00
71 P	Bromoform	0.234	0.238	-1.7	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.319	4.056	6.1	82	0.00
74 T	N-amyl acetate	1.155	1.231	-6.6	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.840	0.834	0.7	86	0.00
76 T	1,2,3-Trichloropropane	0.570	0.639	-12.1	85	0.00
77 T	Bromobenzene	0.940	0.903	3.9	84	0.00
78 T	n-propylbenzene	5.353	5.068	5.3	84	0.00
79 T	2-Chlorotoluene	2.977	2.844	4.5	83	0.00
80 T	1,3,5-Trimethylbenzene	3.546	3.367	5.0	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.311	0.313	-0.6	85	0.00
82 T	4-Chlorotoluene	3.105	2.993	3.6	84	0.00
83 T	tert-Butylbenzene	3.064	2.918	4.8	83	0.00
84 T	1,2,4-Trimethylbenzene	3.515	3.354	4.6	83	0.00
85 T	sec-Butylbenzene	4.584	4.298	6.2	82	0.00
86 T	p-Isopropyltoluene	3.767	3.582	4.9	83	0.00
87 T	1,3-Dichlorobenzene	1.839	1.758	4.4	84	0.00
88 T	1,4-Dichlorobenzene	1.849	1.753	5.2	84	0.00
89 T	n-Butylbenzene	3.709	3.543	4.5	84	0.00
90 T	Hexachloroethane	0.736	0.707	3.9	83	0.00
91 T	1,2-Dichlorobenzene	1.637	1.565	4.4	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.145	0.143	1.4	88	0.00
93 T	1,2,4-Trichlorobenzene	0.995	0.959	3.6	85	0.00
94 T	Hexachlorobutadiene	0.513	0.483	5.8	85	0.00
95 T	Naphthalene	2.114	2.136	-1.0	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011922\
Data File : VY007216.D
Acq On : 19 Jan 2022 10:49
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 20 00:15:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.848	0.1	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6