



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012122\
 Data File : VY007247.D
 Acq On : 21 Jan 2022 09:58
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Title : SW846 8260
 Last Update : Fri Jan 14 05:48:29 2022



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	10676	PASS
75	95	30	60	53.9	29285	PASS
95	95	100	100	100.0	54336	PASS
96	95	5	9	6.7	3634	PASS
173	174	0.00	2	1.4	562	PASS
174	95	50	100	76.4	41533	PASS
175	174	5	9	7.8	3224	PASS
176	174	95	101	96.0	39867	PASS
177	176	5	9	7.6	3015	PASS