

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012397.D
 Acq On : 30 Jan 2023 15:01
 Operator : KP/MD
 Sample : 01358-01
 Misc : 11.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-11-012723

Quant Time: Jan 31 01:48:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	150520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	250055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	242972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	109842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	97838	80.844	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 161.680%	
35) Dibromofluoromethane	7.716	113	75075	56.701	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.400%	
50) Toluene-d8	10.179	98	298716	66.548	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 133.100%	
62) 4-Bromofluorobenzene	12.477	95	97472	52.222	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.440%	
Target Compounds						
52) Toluene	10.240	92	4623	1.069	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012397.D
Acq On : 30 Jan 2023 15:01
Operator : KP/MD
Sample : 01358-01
Misc : 11.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-11-012723

Quant Time: Jan 31 01:48:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

