

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012400.D
 Acq On : 30 Jan 2023 16:11
 Operator : KP/MD
 Sample : 01358-10
 Misc : 4.29g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JC-3-012723

Quant Time: Jan 31 01:48:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	162276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	267680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	264064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	130219	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	104594	80.130	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	160.260%
35) Dibromofluoromethane	7.716	113	80297	56.652	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.300%
50) Toluene-d8	10.179	98	321341	66.898	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	133.800%
62) 4-Bromofluorobenzene	12.477	95	111243	55.676	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012400.D
Acq On : 30 Jan 2023 16:11
Operator : KP/MD
Sample : 01358-10
Misc : 4.29g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JC-3-012723

Quant Time: Jan 31 01:48:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

