

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012402.D
 Acq On : 30 Jan 2023 16:58
 Operator : KP/MD
 Sample : 01354-02
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Jan 31 01:48:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	3098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	5313	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	6699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2698	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	2225	89.775	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	179.540%#
35) Dibromofluoromethane	7.710	113	637	22.643	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	45.280%#
50) Toluene-d8	10.179	98	6677	70.256	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	140.520%#
62) 4-Bromofluorobenzene	12.483	95	2425	61.148	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	122.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012402.D
Acq On : 30 Jan 2023 16:58
Operator : KP/MD
Sample : 01354-02
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Jan 31 01:48:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

