

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
 Data File : VY012406.D
 Acq On : 30 Jan 2023 18:32
 Operator : KP/MD
 Sample : 01356-01
 Misc : 4.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-J

Quant Time: Jan 31 01:49:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	159808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	269431	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	262992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	129281	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	97158	75.340	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.680%
35) Dibromofluoromethane	7.716	113	78288	54.876	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.760%
50) Toluene-d8	10.179	98	318544	65.813	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	131.620%
62) 4-Bromofluorobenzene	12.483	95	108307	53.854	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY013023\
Data File : VY012406.D
Acq On : 30 Jan 2023 18:32
Operator : KP/MD
Sample : 01356-01
Misc : 4.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-J

Quant Time: Jan 31 01:49:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

